

CERTIFICATE OF COMPLIANCE

Certificate Number 20161220-E480011
Report Reference E480011-20161209
Issue Date 2016-DECEMBER-20

Issued to: HONEYWELL ANALYTICS ASIA PACIFIC CO LTD
7F SangAm IT Tower
434 World Cup Buk-ro Mapo-gu
Seoul 03922 KOREA

**This is to certify that
representative samples of**

GAS AND VAPOR DETECTION EQUIPMENT
CLASSIFIED FOR USE IN HAZARDOUS LOCATIONS;
GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN
ZONE CLASSIFIED HAZARDOUS LOCATIONS

USC, CNC - Class I, Groups A, B, C, and D Hazardous
Locations

USC - Class I, Zone 0, AEx ia IIC Hazardous Locations

CNC - Class I, Zone 0, Ex ia IIC Hazardous Locations

Model BW Clip4 Multi-Gas Detector, Part No. BWC4,
followed by -Y or -B, followed by -N, intrinsically safe when
used with two Energizer L91 batteries.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:
Additional Information:

See Addendum Page

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations>



CERTIFICATE OF COMPLIANCE

Certificate Number 20161220-E480011
Report Reference E480011-20161209
Issue Date 2016-DECEMBER-20

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Standard No. UL 913, Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations, Standard No. UL 913, Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations, Standard No. UL 60079-11, Explosive Atmospheres - Part 11: Equipment Protection by Intrinsic Safety "i", Standard No. CAN/CSA C22.2 No. 152-M1986, Combustible Gas Detection Instruments, Standard No. CAN/CSA C22.2 No. 157-92, Intrinsically Safe and Non-incendive Equipment for Use in Hazardous Locations, Standard No. CAN/CSA-C22.2 No. 60079-0:15, Explosive atmospheres -- Part 0: Equipment -- General requirements, Standard No. CAN/CSA-C22.2 No. 60079-11:14, Explosive atmospheres -- Part 11: Equipment protection by intrinsic safety "i"



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations>

